

TrichoScope Polarizer HR (MEDL7HM)


Merk: Dino-Lite Medical
Productcode: MEDL7HM

Korte omschrijving

5 Megapixel resolution
 Up to 200x magnification
 HR model for Trichoscopy
 Polarizer
 Medical Device Class 1

Omschrijving

The Dino-Lite TrichoScope Polarizer HR (MEDL7HM) is a great instrument for the examination of the human scalp and for hair analysis, e.g. for hair loss research. With the high magnification of up to 200 times a single hair can be viewed with great detail. The built-in polarizer filter minimizes the glaring effect of the scalp/ hair.



The 5 Megapixel sensor makes it possible to display microscopic images, without loss of quality, on large screens or when large images should be printed.

Specificatie

Verlichting	
Licht / LED-type	White
Aantal leds	8
LED aan / uit schakelbaar:	Yes
Infrarood filter	IR cut-filter >650 nm
Diffusor beschikbaar	No
Polarisator	Yes, linear
Optiek	

Magnification	10-70x, 200x
Type lens	Glass with anti-reflection coating
Sensor	
Sensortype	CMOS
Resolutie	5 Megapixel (2592x1944)
Maximale framesnelheid	30 fps
Compatibiliteit	
Interface	USB 2.0
Besturingssysteem	Windows 7, 8, 10 & 11, MacOS 10.9 and up
Software	DinoCapture 2.0 (Windows), DinoXcope (Mac OS)
Ondersteunde beeldformaten (Windows)	BMP, GIF, PNG, JPG, TIF, RAS, PNM, TGA, PCX, MNG, WBMP, JP2, JPC, PGX
Ondersteunde videoformaten (Windows)	WMV, FLV, SWF
Ondersteunde beeldformaten (MacOS)	JPEG, PNG
Ondersteunde videoformaten (MacOS)	MOV (max 1.3MP)
Beeldvormingsnormen	DirectShow, UVC
Behuizing	
Materiaal van de behuizing	Aluminium alloy housing
Vergrotingsslot	No
Afmetingen	10.5cm (L) x 3.2cm (D)
Gewicht	140g
Kabellengte	1.8m
Kenmerken	
Bijzonderheid	No
Meting	No
Kalibratie	No
Microtouch-sensor	Yes
Informatie	
Inhoud van de verpakking	Microscope, carry pouch, software CD, ZT-Z-CC1 cap closure, user manual
Garantie-informatie	2 years European warranty
Regelgevende goedkeuring	Medical Device Class 1 – Medical Devices Regulation (EU) 2017/745
Prijsklasse	€700,00 - €900,00

Product galerij

